

Rua Luiz da Costa Ramos, 260 - Saúde CEP 04157-020 - São Paulo - SP - Brasil Tel.: (11) 3056-1900 - Fax: (11) 5073-3366 E-mail: vendas@presys.com.br Site: www.presys.com.br







EC Declaration of Conformity

We declare under our sole responsability that the CE marked products, are in conformity with the essential requirements of the following EC Directives when installed in accordance with the installation instructions contained in the product documentation:

Series	TE-25N, TE-35N and TE-50N	
Description	Dry-Block Temperature Calibrator	

LVD Low Voltage Directive	2014/35/EC of the European Parliament and of the Council of 12 December 2006 on the harmonization of the laws of Member States relating to Electrical Equipment designed for use within certain voltage limits.		
EN 61010-1:2011	Safety requirements for electrical equipment for measurement, control and laboratory use		
EN 61010-2:010	Safety requirements for electrical equipment for measurement, control and laboratory use - Part 2-010: Particular requirements for laboratory equipment for the heating of Materials.		
EMC directive	2004/108/EC of the European Parliament and of the Council of 15 December 2004 on the approximation of the laws of the Member States relating to electromagnetic compatibility and repealing Directive 89/336/EEC		
EN 61326-1:2003	Electrical equipment for measurement, control and laboratory use - EMC requirements		

São Paulo, 8 September 2015

Vii	Ref.
Vinicius José Gomes Nunes	Antonio Rafael Sito Antunes
CEO	Engineering Manager